

INTENDED USE

CBC-3K is an assayed whole blood control designed to monitor values on multi parameter hematology cell counter. Please refer to the assay table for specific instrument models.

SUMMARY AND PRINCIPLE

It is an established laboratory practice to use a stable control to monitor the performance of diagnostic tests. This control is composed of stable materials that provide a means of monitoring the performance of hematology blood cell counters. It is sampled in the same manner as a patient specimen.

REAGENTS

CBC-3K is an *in vitro* diagnostic reagent composed of human erythrocytes, mammalian leukocytes, and mammalian platelets suspended in a plasma-like fluid with preservatives.

**PRECAUTION**

CBC-3K is intended for *in vitro* diagnostic use only by trained personnel.

**WARNING:**

POTENTIAL BIOHAZARDOUS MATERIAL. For *in vitro* diagnostic use. Each human donor/unit used in the preparation of this product has been tested and yielded non-reactive / negative results for all conditions referenced in 21 CFR 610.40 (a) (b), as required by the FDA. Testing was conducted using FDA-licensed tests. Additional details can be found at:

<http://www.rndheme.com/TechnicalInformation.aspx>

No test method can offer complete assurance that infectious agents are absent; therefore, this material should be handled as potentially infectious. When handling or disposing of vials follow precautions for patient specimens as specified in the OSHA Bloodborne Pathogen Rule (29 CFR Part 1910, 1030) or other equivalent biosafety procedures.

**STABILITY AND STORAGE**

Store CBC-3K upright at 2 - 8° C (35 - 46° F) when not in use.

Protect tubes from overheating and freezing. Unopened tubes are stable through the expiration date. Opened product is stable for **8 or 15 days (depending to your analyzer, refer to the assay sheet) or 15 pierces**, whichever comes first, provided they are handled properly.

INDICATIONS OF DETERIORATION

After mixing, product should be similar in appearance to fresh whole blood. In unmixed tubes, the supernatant may appear cloudy and reddish; this is normal and does not indicate deterioration. Other discoloration, very dark red supernatant or unacceptable results may indicate deterioration. **Do not use the product if deterioration is suspected.**

**INSTRUCTIONS FOR USE**

1. Remove tubes from the refrigerator and allow to warm to room temperature (15 to 30°C or 59 to 86°F) for 15 minutes before mixing.
2. To mix, hold a tube horizontally between the palms of the hands. **Do not pre-mix on a mechanical mixer.**
 - a) Roll the tube back and forth for 20 - 30 seconds; occasionally invert the tube. Mix vigorously, but do not shake.
 - b) Continue to mix in this manner until the red cells are completely suspended. Tubes stored for a long time may require extra mixing.

- c) Gently invert the tube 8 - 10 times immediately before sampling.

3. Analyze the sample as instructed in the Quality Control section of the Operator's Manual for your instrument.
4. After sampling:
 - a) If tube has been opened for sampling, clean residual material from the cap and tube rim with a lint-free tissue. Replace the cap tightly.
 - b) Return tubes to refrigerator within 30 minutes of use.

EXPECTED RESULTS

Verify that the lot number on the tube matches the lot number on the table of assay values. Assay values are determined on well-maintained, properly calibrated instruments using the instrument manufacturer's recommended reagents. Reagent differences, maintenance, operating technique, and calibration may contribute to inter-laboratory variation.

PERFORMANCE CHARACTERISTICS

Assigned values are presented as a Mean and Range. The Mean is derived from replicate testing on instruments operated and maintained according to the manufacturer's instructions. The Range is an estimate of variation between laboratories and also takes into account inherent imprecision of the method and expected biological variability of the control material.

Assay values on a new lot of control should be confirmed before the new lot is put into routine use. Test the new lot when the instrument is in good working order and quality control results on the old lot are acceptable. The laboratory's recovered mean should be within the assay range.

For greater control sensitivity each laboratory should establish its own mean and acceptable range and periodically reevaluate the mean. The laboratory range may include values outside of the assay range. The user may establish assay values not listed on the Assay Sheet, if the control is suitable for the method.

LIMITATIONS

The performance of this product is assured only if it is properly stored and used as described in this insert. Incomplete mixing of a tube prior to use invalidates both the sample withdrawn and any remaining material in the tube.

TECHNICAL ASSISTANCE AND CUSTOMER SERVICE

For technical assistance or additional information, please call your dealer or local distributor. If there is no, you may call Bio-techne® Technical Service at 33 (0)2 99 35 19 36.

QUALITY CONTROL PROGRAM

For information on the Inter-Laboratory Quality Control Program, please call Bio-techne® CBC-Monitor Service at 33 (0)2 99 35 19 36.

All brands and products are trademarks or registered trademarks of their respective companies.

**R & D Systems, Inc.**614 McKinley Place NE
Minneapolis, MN USA 55413

IS033-016 Rev 08/17

**Bio-techne®**19 Rue Louis Delormel
35230 Noyal Châtillon / Seiche
France

IS03K00-V12 10/2019



CBC-3K®

Chất kiểm soát huyết học (Hematology Control)

Mục đích sử dụng (INTENDED USE)

- CBC-3K là **chất kiểm soát máu toàn phần có định giá**, dùng để giám sát giá trị trên các máy phân tích huyết học **đa thông số**.
- Vui lòng tham khảo **bảng định giá (assay table)** cho từng dòng thiết bị cụ thể

IS03K00 Instruction sheet CBC-3...

.

Tóm tắt và nguyên lý (SUMMARY AND PRINCIPLE)

- Trong phòng xét nghiệm, việc sử dụng chất kiểm soát ổn định để theo dõi hiệu năng của xét nghiệm chẩn đoán là thông lệ chuẩn.
- CBC-3K được cấu tạo từ các thành phần ổn định, giúp giám sát hiệu năng của máy đếm tế bào máu.
- Được lấy mẫu tương tự như mẫu bệnh nhân

IS03K00 Instruction sheet CBC-3...

.

Thành phần (REAGENTS)

CBC-3K là thuốc thử chẩn đoán **in vitro**, gồm:

- Hồng cầu người,
- Bạch cầu động vật có vú,
- Tiêu cầu động vật có vú,
- Treo trong dung dịch tương tự huyết tương có chất bảo quản

IS03K00 Instruction sheet CBC-3...

.

Thận trọng (PRECAUTION)

- Chỉ sử dụng cho chẩn đoán **in vitro**, bởi nhân viên đã qua đào tạo.
- **CẢNH BÁO: Nguy cơ sinh học tiềm tàng.**
 - Mỗi đơn vị máu người dùng trong sản xuất đã được kiểm tra theo **21 CFR 610.40 (a)(b)** bằng test FDA phê duyệt → kết quả âm tính/không phản ứng.
 - Tuy nhiên, không có phương pháp nào đảm bảo loại trừ hoàn toàn mầm bệnh → phải xử lý như **mẫu bệnh phẩm có khả năng lây nhiễm**, tuân thủ OSHA Bloodborne Pathogen Rule (29 CFR 1910.1030) hoặc quy trình an toàn sinh học tương đương

IS03K00 Instruction sheet CBC-3...

Ôn định và bảo quản (STABILITY AND STORAGE)

- Bảo quản **thẳng đứng** ở **2 – 8°C (35 – 46°F)**.
- Tránh quá nóng hoặc đông lạnh.
- Ống chưa mở: ôn định đến ngày hết hạn ghi trên nhãn.
- Ống đã mở: ôn định **8 – 15 ngày** (tùy loại máy, xem trong bảng QC) hoặc tối đa **15 lần chọc kim**, tùy điều kiện nào đến trước

IS03K00 Instruction sheet CBC-3...

Dấu hiệu hư hỏng (INDICATIONS OF DETERIORATION)

- Sau khi trộn: sản phẩm phải có hình thái giống máu toàn phần tươi.
- Ống chưa trộn: dịch nổi có thể đục, hơi đờ → bình thường.
- Nếu dịch nổi đờ sẫm/nâu hoặc kết quả QC không đạt → nghi ngờ hư hỏng → **không sử dụng**

IS03K00 Instruction sheet CBC-3...

Hướng dẫn sử dụng (INSTRUCTIONS FOR USE)

1. Lấy ống ra khỏi tủ lạnh, để ở nhiệt độ phòng (15 – 30°C) trong 15 phút.
2. Trộn:
 - o Giữ ống ngang giữa hai lòng bàn tay, lăn qua lại 20 – 30 giây, thỉnh thoảng đảo ngược.
 - o Không dùng máy lắc cơ học.
 - o Trộn mạnh nhưng không lắc dốc.
 - o Ống để lâu có thể cần trộn kỹ hơn.
3. Trước khi lấy mẫu: đảo ngược 8 – 10 lần.
4. Phân tích theo phần **Quality Control** trong **Operator's Manual** của máy.
5. Sau khi lấy mẫu:
 - o Lau sạch nắp và miệng ống bằng giấy không xơ.
 - o Đậy kín nắp.
 - o Trả lại vào tủ lạnh trong vòng 30 phút

IS03K00 Instruction sheet CBC-3...

Kết quả mong đợi (EXPECTED RESULTS)

- Đảm bảo số lô trên ống trùng với bảng QC.
- Giá trị QC xác lập trên máy hiệu chuẩn, bảo dưỡng tốt, dùng thuốc thử khuyến cáo.
- Khác biệt thuốc thử, thao tác, hiệu chuẩn, bảo dưỡng → có thể gây biến thiên liên phòng

IS03K00 Instruction sheet CBC-3...

Đặc tính hiệu năng (PERFORMANCE CHARACTERISTICS)

- Giá trị QC được trình bày dưới dạng **Trung bình (Mean)** và **Khoảng (Range)**.
- Mean: từ nhiều lần lặp lại trên máy hiệu chuẩn đúng.
- Range: ước tính biến thiên liên phòng, gồm sai số phương pháp và biến thiên sinh học.
- Khi có lô mới: phải xác nhận QC trước khi dùng thường quy.
- Mỗi phòng xét nghiệm có thể thiết lập **mean và range riêng**, có thể rộng hơn hoặc khác so với bảng QC hàng

Giới hạn (LIMITATIONS)

- Hiệu năng chỉ đảm bảo khi bảo quản và sử dụng đúng hướng dẫn.
- Nếu không trộn đầy đủ trước khi dùng → cả mẫu lấy ra và phần còn lại trong ống đều **không hợp lệ**

Hỗ trợ kỹ thuật (TECHNICAL ASSISTANCE)

- Liên hệ nhà phân phối địa phương.
- Nếu không, gọi **Bio-techne® Technical Service: +33 (0)2 99 35 19 36**

Chương trình kiểm soát chất lượng (QUALITY CONTROL PROGRAM)

Thông tin về **Chương trình QC liên phòng (Inter-Laboratory Quality Control Program)** có tại dịch vụ **CBC-Monitor** của **Bio-techne®**